



Ferticare 2.0

www.MyPelvicHealth.co.uk



PENILE VIBRATORY SYSTEM

indicated for anejaculation (S.C.I. above T10) and delayed ejaculation to help achieve ejaculation

Frequency Range : 70-110 Hz

Amplitude Range (peak to peak): 0.5 – 4.0 mm Variable

One Year warranty is included with purchase (valid for original owner only)

Made in the U.S.A. by Reflexonic LLC

This item is not returnable or refundable when used